



Dongguan Yuanyao Electronics Technology Co., Ltd.

Add.: No.3, 3rd Street, Yinsha Road, Xiabian Community, Chang'an Town, Dongguan city, Guangdong, P.R.China

Tel: +86-769-86245189

Fax: +86-769-86245990

Vacuum Oven Test Report

Customer: Sateco Labor

Model: YPOZ-072

Production No.: YY-20180621

Temperature: RT+10°C~200°C

Specification					
Temperature test					
Set temperature	Temperature		Time		Date 日期
	Controller	Actual	Controller time	Inside temperature arriving time	
100°C	100°C	100°C	20 min	26 min	2018/09/3
150°C	150°C	150°C	30 min	40 min	2018/09/3
180°C	180°C	180°C	45 min	50 min	2018/09/4
200°C	200°C	200°C	60 min	65 min	2018/09/4
<p style="text-align: center;">Set temperature: 150°C</p> <p style="text-align: center;">A: 153.5 B:152.3 C:153.7 D:156.4 O:151.6 K: 144.8 L: 135.4 N:136.8 M: 146.4</p> <p>Temperature uniformity: ±12.5°C</p> <p style="text-align: center;">Set temperature: 200°C</p> <p style="text-align: center;">A: 204.8 B:202.0 C:205.2 D:205.6 O:202.1 K: 198.5 L: 180.6 N:181.3 M: 194.6</p> <p>Temperature uniformity: ±10.5 °C</p>					
Vacuum test					
Set Vacuum	Actual vacuum		Time	Date	
-30Kpa	-30.66 Kpa		36 S	2018/09/3	
-50Kpa	-51.99 Kpa		41 S	2018/09/3	
-98Kpa	-98.32 Kpa		123 S	2018/09/3	
-101Kpa	-100.99 Kpa		155 S	2018/09/3	



Maim part specification		
Item	Specification	Date
Temperature heater	1.5KW*3	2018/09/3
Vacuum pump	Leybold SV 16B	2018/09/3
Vacuum oil	LVO 120	2018/09/3
PU is filling good?	Yes	2018/09/15
Appearance of chamber is clear?	Yes	2018/09/15
Instruction:warning labels pasted correct?	Yes	2018/09/15

For and on behalf of
Dongguan Yuanyao Electronics Technology Co., Ltd.

Tel: +86-769-86245189 Ext.810 | Fax: +86-769-86245990 | Mobile: +86-15876432270

E-mail: sales06@yuanyao-tech.com | Wechat/whatsapp: +86 15876432270

Web: www.gdmypt.com Showroom on alibaba.com: www.my-pt.en.alibaba.com



Room pressure, the temperature date after 80 minutes stable

Note: Vacuum oven is heated by heat radiation. This need more time to be stable. For and on behalf of Dongfang Technology Co.,Ltd.

Approved: Dongfang Huang

Inspector: Zongjing Zuo

Authorized Signature(s)
2018-9-12

Tel: +86-769-86245189 Ext.810 | Fax: +86-769-86245990 | Mobile: +86-15876432270

E-mail: sales06@yuanyao-tech.com | Wechat/whatsapp: +86 15876432270

Web: www.gdmypt.com Showroom on alibaba.com: www.my-pt.en.alibaba.com